



CERTIFICATE OF CALIBRATION
 ISSUED BY YOUNG CALIBRATION LIMITED
 DATE OF ISSUE: 04/04/17
 CERTIFICATE NUMBER: C53183



Young Calibration Ltd
 5 Cecil Pashley Way
 Shoreham Airport
 Shoreham-by-Sea
 West Sussex, BN43 5FF

+44 (0)1273 455572
 enquiries@youngcalibration.co.uk
 www.youngcalibration.co.uk

Document : YCF/023
 Issue Number : 2.9
 Page 1 of 1 Page
 Approved Signatory

A. Young M. Hindle
 N. Mardon C. Millard

Customer: Fasinternational Srl	Manufacturer & Model: MKS 1179
Customer Address: Via Selvanesco, 75/77	UUT Description: Mass Flow Controller
20142 Milano (MI) Italy	Customer ID Number: n/a
For and on behalf of: Consorzio L.E.A.P	Serial Number: G505987G20
Address: Via Nino Bixio 27/c	Equipment Condition: Used / As Received / Final
29121, Piacenza	Nominal Calibration Range: 0 to 5000 scc/min
Purchase Order Number: 201700332 (424)	Calibration Fluid: Dry Air
Date of Receipt: 29 March 2017	Calibration Fluid Temperature: 18.2 to 20.3 °C
YCL Project Number: YC/39140	Calibration Location: YCL Laboratory
Calibration Date: 04 April 2017	Laboratory Temperature: 18.7 to 20.7 °C
Requested Due Date: n/a	Laboratory Barometric Pressure: 1024.8 to 1026.0 mbarA
Calibration Performed By: C. Millard	Laboratory Relative Humidity: 57.1 to 59.4 %RH
Calibration Procedure: Procedure 530	Reference Equipment: YC/010/211 to 217

Calibration Method & Notes

The UUT (unit under test) was mounted in the Young Calibration Low Gas Flow Calibration Rig and was calibrated using a series transfer method. Clean, dry air was blown through the UUT to the desired flowrate.

When stabilised conditions were observed, the measurements conditions were recorded. The flowrate was adjusted to the next condition, and once steady state conditions were achieved, the results were again recorded, this procedure being repeated until the calibration was complete. The UUT results are derived from the average of at least 9 readings.

The standard flowrate is referenced to conditions of 1013.25 mbarA and 0 °C

Calibration Results

Standard Flowrate (scc/min)	UUT Output (V)	UUT Output Uncertainty (± V)
0.00	0.000	0.007
1015.86	1.008	0.007
2007.91	1.993	0.007
3024.43	3.020	0.007
4036.38	4.015	0.007
5023.78	4.997	0.007

Temperature of gas at the UUT : 18.3 - 20.2 °C
 Pressure at the UUT : 1025.4 - 1025.5 mbarA

The uncertainty of the above flowrate measurements under laboratory conditions is ± 1.0 % (+ UUT Uncertainty)

The uncertainties reported refer to the measured values only and not to the ability of the instrument to maintain its calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. UKAS is one of the signatories to the Multilateral Agreement of the European co-operation for Accreditation (EA) for the mutual recognition of calibration certificates issued by accredited laboratories.

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes.

The certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.