

Nippon Gases Belgium NV
Niiverheidsstraat 4
B-2260 Oevel - België
Tel : +32(0)14/250.611
Fax : +32(0)14/581.505

B9016235
4565E/12267078



Date : 08/05/19
Praxair Ref. Nr. : 398,531,001
Customer Ref. Nr. : 2805 OS

To : RIVOIRA GAS SRL(SPG)
VIA MARIE CURIE 134/D

10034 CHIVASSO (TORINO)
ITALIE

Cylinder Number : 12267078

Cylinder Type : 10H
Valve Connection : N14 HPBR

Cylinder Pressure : 150.0 bara
Cylinder Content : 1500 Liter

Page 1/1

CERTIFICATE OF ANALYSIS : CALIBRATION MIXTURE

CERTIFIED GAS MIXTURE

COMPONENTS	REQUESTED	RESULTS	UNCERTAINTY (%)
PROPANE	10.0 MOLPPM	9.54 MOLPPM	± 1 REL
METHANE	25.0 MOLPPM	25.94 MOLPPM	± 1 REL
AIR	BALANCE	BALANCE	

ANALYST :

GREET VERWAEST

We guarantee the stability of the product until 11/05/24 as far as it has not received an improper treatment.

Remarks :

- Do not use/store below -20 °C or above +50°C.
- The reported uncertainty is based on the standard uncertainty multiplied with a coverage factor $k=2$ which corresponds to a confidence interval of approximately 95%.
- The reported results are traceable to (inter)national standards.