

Certificate of Calibration and Conformance

Certificate Number 2011-146015

Instrument Model 831, Serial Number 0002631, was calibrated on 11JUL2011. The instrument meets factory specifications per Procedure D0001.8310, ANSI S1.4-1983 (R 2006) Type 1; S1.4A-1985 ; S1.43-1997 Type 1; S1.11-2004 Octave Band Class 0; S1.25-1991; IEC 61672-2002 Class 1; 60651-2001 Type 1; 60804-2000 Type 1; 61260-2001 Class 0; 61252-2002.

New Instrument
Date Calibrated: 11JUL2011
Calibration due:

Calibration Standards Used

MANUFACTURER	MODEL	SERIAL NUMBER	INTERVAL	CAL. DUE	TRACEABILITY NO.
Stanford Research Systems	DS360	61889	12 Months	01FEB2012	61889-020111

Reference Standards are traceable to the National Institute of Standards and Technology (NIST)

Calibration Environmental Conditions

Temperature: 24 ° Centigrade

Relative Humidity: 42 %

Affirmations

This Certificate attests that this instrument has been calibrated under the stated conditions with Measurement and Test Equipment (M&TE) Standards traceable to the U.S. National Institute of Standards and Technology (NIST). All of the Measurement Standards have been calibrated to their manufacturers' specified accuracy / uncertainty. Evidence of traceability and accuracy is on file at Provo Engineering & Manufacturing Center. An acceptable accuracy ratio between the Standard(s) and the item calibrated has been maintained. This instrument meets or exceeds the manufacturer's published specification unless noted.

This calibration complies with the requirements of ISO 17025 and ANSI Z540. The collective uncertainty of the Measurement Standard used does not exceed 25% of the applicable tolerance for each characteristic calibrated unless otherwise noted.

The results documented in this certificate relate only to the item(s) calibrated or tested. A one year calibration is recommended, however calibration interval assignment and adjustment are the responsibility of the end user. This certificate may not be reproduced, except in full, without the written approval of the issuer.

Tested with PRM831-019146

Signed: 
Technician: Ron Harris

Certificate of Calibration and Conformance

Certificate Number 2011-145987

Instrument Model PRM831, Serial Number 019146, was calibrated on 11JUL2011. The instrument meets factory specifications per Procedure D0001.8167.

New Instrument
Date Calibrated: 11JUL2011
Calibration due:

Calibration Standards Used

MANUFACTURER	MODEL	SERIAL NUMBER	INTERVAL	CAL. DUE	TRACEABILITY NO.
Hewlett Packard	34401A	MY41044529	12 Months	26JAN2012	5056765
Larson Davis	LDSigGn/2209	0277 / 0109	12 Months	21MAR2012	2011-141059

Reference Standards are traceable to the National Institute of Standards and Technology (NIST)

Calibration Environmental Conditions

Temperature: 24 ° Centigrade

Relative Humidity: 42 %

Affirmations

This Certificate attests that this instrument has been calibrated under the stated conditions with Measurement and Test Equipment (M&TE) Standards traceable to the U.S. National Institute of Standards and Technology (NIST). All of the Measurement Standards have been calibrated to their manufacturers' specified accuracy / uncertainty. Evidence of traceability and accuracy is on file at Provo Engineering & Manufacturing Center. An acceptable accuracy ratio between the Standard(s) and the item calibrated has been maintained. This instrument meets or exceeds the manufacturer's published specification unless noted.

This calibration complies with the requirements of ISO 17025 and ANSI Z540. The collective uncertainty of the Measurement Standard used does not exceed 25% of the applicable tolerance for each characteristic calibrated unless otherwise noted.

The results documented in this certificate relate only to the item(s) calibrated or tested. A one year calibration is recommended, however calibration interval assignment and adjustment are the responsibility of the end user. This certificate may not be reproduced, except in full, without the written approval of the issuer.

Signed: 
Technician: Ron Harris

Certificate of Calibration and Conformance

Certificate Number 2011-145666

Microphone Model 377B02, Serial Number 125997, was calibrated on 01JUL2011. The microphone meets factory specifications per Test Procedure D0001.8167.

New Instrument
Date Calibrated: 01JUL2011
Calibration due:

Calibration Standards Used

MANUFACTURER	MODEL	SERIAL NUMBER	INTERVAL	CAL. DUE	TRACEABILITY NO.
Larson Davis	PRM915	0102	12 Months	17AUG2011	2010-132962
Larson Davis	PRM902	0206	12 Months	17AUG2011	2010-132963
Larson Davis	2559	3034LF	12 Months	18AUG2011	2010-133036
Larson Davis	PRM902	0529	12 Months	08SEP2011	2010-133837
Larson Davis	PRM902	0528	12 Months	08SEP2011	2010-133838
Larson Davis	MTS1000 / 2201	1000 / 0100	12 Months	10SEP2011	SM090910
Hewlett Packard	34401A	3146A62099	12 Months	11NOV2011	4994123
Larson Davis	2559	2504	12 Months	29NOV2011	17865-1
Larson Davis	PRM916	0102	12 Months	23DEC2011	2010-137908
Larson Davis	CAL250	42630	12 Months	04JAN2012	2011-138110
Larson Davis	2900	0575	12 Months	14JUN2012	2011-144882

Reference Standards are traceable to the National Institute of Standards and Technology (NIST)

Calibration Environmental Conditions

Environmental test conditions as printed on microphone calibration chart.

Affirmations

This Certificate attests that this instrument has been calibrated under the stated conditions with Measurement and Test Equipment (M&TE) Standards traceable to the U.S. National Institute of Standards and Technology (NIST). All of the Measurement Standards have been calibrated to their manufacturers' specified accuracy / uncertainty. Evidence of traceability and accuracy is on file at Provo Engineering & Manufacturing Center. An acceptable accuracy ratio between the Standard(s) and the item calibrated has been maintained. This instrument meets or exceeds the manufacturer's published specification unless noted.

This calibration complies with the requirements of ISO 17025 and ANSI Z540. The collective uncertainty of the Measurement Standard used does not exceed 25% of the applicable tolerance for each characteristic calibrated unless otherwise noted.

The results documented in this certificate relate only to the item(s) calibrated or tested. A one year calibration is recommended, however calibration interval assignment and adjustment are the responsibility of the end user. This certificate may not be reproduced, except in full, without the written approval of the issuer.

Signed: Abraham Ortega
Technician: Abraham Ortega